

 <p>IIT PALAKKAD</p>	<p>भारतीय प्रौद्योगिकी संस्थान पालक्काड Indian Institute of Technology Palakkad अहलिआ एकीकृत कैम्पस, कोज़िहपारा Ahalia Integrated Campus, Kozhipara पालक्काड- 678557 Palakkad – 678 557</p>	<p>Phone no: 04923 – 226 300 /586 Email : purchase@iitpkd.ac.in</p>
---	--	--

REF NO: IITPKD/CFF/RP/093/2017

DATE: 03.01.2018

DUE DATE OF THE TENDER : 17.01.2018 AT 3.00 PM

CORRIGENDUM

Clarifications for IITPKD/CFF/RP/093/2017, Date: 27.12.2017

1. Technical Specification, Point 5: [Alignment stage]
The X and Y movements of the alignment stage should be motorised.
2. Technical Specification, Point 7: [Top-side microscope]
Clear explanations of how the optical magnification is atleast 450X, or higher, must be provided.
3. Technical Specification, Point 8: [Bottomside microscope]
The optical magnification of the bottomside microscope should be same, or higher, than the top-side microscope ($\geq 450X$ optical).

Changes for IITPKD/CFF/RP/093/2017, Date: 27.12.2017

1. Technical Specification, Point 3: [Lamp housing]
 - A. Point 3.a.ii: 1000 W Hg lamp is an **optional** item. Price must be quoted separately.
 - B. Point 3.h: **Only one** extra lamp for each of the specified lamp wattages/type is required.
2. Technical Specification, Point 12: [System control]
The vendor is to provide the computer (minimum specification 8GB RAM & Intel i5 processor & 1TB HDD) along with all the required software specified in Point 12.

Rest of the terms and conditions of the tender will remain unchanged

REGISTRAR, IIT PALAKKAD