


**NOTICE INVITING TENDER (NIT) FOR SUPPLY, INSTALLATION, TESTING  
AND COMMISSIONING OF THERMO VACUUM CHAMBER**

**Tender No. IITPKD/MECH/SC/036/2019-20**

**Date/Time of Publication: 19-08-2019, 1500 hrs**

**Date/Time of Pre-bid Meeting: 27-08-2019, 1500 hrs**

**Date/Time of Closing: 02-09-2019, 1500 hrs**



**IIT PALAKKAD**

**Indian Institute of Technology Palakkad  
Ahalia Integrated Campus, Kozhipara,  
Palakkad – 678 557.**

## 1. GENERAL

1.1. Indian Institute of Technology Palakkad (herein after called “IITPKD”) invites sealed Tenders under Two-Bid System for **Supply, Installation, Testing and Commissioning of Thermo Vacuum Chamber** as per the technical specifications given in **Annexure-1.2**. The tender documents may be obtained from the **Office of the Registrar, IIT Palakkad, Ahalia Integrated Campus, Kozhipara, Palakkad-678557** or downloaded from <https://iitpkd.ac.in/tenders> or <https://eprocure.gov.in/cppp/>. Last date/time for submission of the bids is **02-09-2019, 1500 hrs**. The bids will be opened by the duly constituted Committee in the presence of the bidders or their authorized representatives, who wish to be present on the same day at **02-09-2019, 1515 hrs**. The bidder’s representative should carry authorization letter from their company empowering them to participate in the pre-bid and tender opening meetings. The technical bids will be opened and examined by a technical committee, which will decide the suitability of the bid as per the specifications and requirements of IITPKD. In respect of opening of financial bids, those bidders who are technically qualified only will be called. In case of any holiday or unforeseen closure of the institute on the scheduled day of opening of the bids, the bids will be opened on the next working day at the same time, but the deadline for submission of bids remains the same as indicated above. Supply/Download of tender documents will close on the last working day at 1500 hours before the last date for submission of the tenders.

1.3. The bids shall reach **Registrar, IIT Palakkad, Ahalia Integrated Campus, Kozhipara, Palakkad-678557**, by Post/Courier/in person latest by **02-09-2019, 1500 hrs**. Bids received after the above mentioned date and time shall not be considered. Bids sent through Cable/Facsimile/Email/FAX/any other mode shall not be considered. Conditional bids will be rejected outright. The tender box is kept in the office of the Academic Block, IIT Palakkad, Ahalia Integrated Campus, Kozhipara, Palakkad-678 557.

1.4. The responsibility of submission of the bids on or before the last date shall rest with the tenderer. The institute will hold no responsibility for the non-receipt of the bids or for the bids received after the date/time specified. Any bid received by IITPKD after the bid submission deadline prescribed by IITPKD, shall be rejected and returned unopened to the Bidder.

1.5. The timeline for the NIT is as mentioned below:

S. No.	Events	Date and Time
1	<b>Publication of the Tender Document</b>	<b>19-08-2019, 1500 hrs</b>
2.	<b>Pre-bid Meeting</b>	<b>27-08-2019, 1500 hrs</b>
3.	<b>Last date for submission of sealed tenders</b>	<b>02-09-2019, 1500 hrs</b>
4	<b>Opening of Technical Bids</b>	<b>02-09-2019, 1515 hrs</b>

1.6. Canvassing or offer of an advantage or any other inducement by any person with a view to influencing acceptance of a bid is an offence under Laws of India. Such action will result in the rejection of bid, in addition to other punitive measures.

1.7. Each tenderer shall submit only one bid, either by himself or as a partner in a joint venture or as a member of consortium. If a bidder or if any of the partners in a joint venture or any one of the members of the consortium participate in more than one bid, the bids (of both the individual and the partnership/consortium/joint venture) are liable to be rejected.

1.8. The bidder shall bear all costs associated with the preparation and submission of his bid and IITPKD shall in no case be responsible or liable for those costs, regardless of the conduct or outcome of the tender process.

1.9. The Tender Document is not transferable. The bidder shall make a copy of the tender document before submitting the same to the concerned office. No requests will be entertained for making a copy after the submission of the document.

1.10. IITPKD will respond to any request for clarification or modification of the Tender Document that are received up to **FIVE (05) days** prior to the deadline for submission of bids

prescribed by IITPKD. For this purpose, the prospective bidder(s) requiring clarification in the Tender Document shall notify IITPKD in writing at the address mentioned. Any such clarification, together with all details on which the clarification had been sought, will be published in the website.

1.11. Except for any such written clarification by the Institute, which is expressly stated to be an addendum to the tender document issued by the Registrar, IIT Palakkad, no written or oral communication, presentation or explanation by any other employee of any of the Sections/Departments of the Institute, shall be taken to bind or fetter the Institute.

## **2. AMENDMENTS IN THE TENDER DOCUMENT**

2.1. At any time prior to the deadline for submission of bids, IITPKD may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective Bidder, modify the Tender Document by way of amendment(s).

2.2. Amendments will be intimated through the institute's website and the tenderers shall ensure that the amendments are carried out in the bid before submission. The amendments will not be published in newspapers. Bidders should regularly visit the institute's website to keep themselves updated.

2.3. No extension in the bid due date/ time shall be considered on account of delay in receipt of any document by mail. Further, it will be assumed that the Bidder has taken into account, such amendments, while submitting the bid.

## **3. COMPOSITION OF THE TENDER DOCUMENT**

3.1. The Tender Document comprises of:

- (a) Instruction to the bidders including terms and conditions
- (b) Technical Specifications (Annexure-I)
- (c) Pre-qualification Criteria (Annexure-II)
- (d) Techno-Commercial Parameters (Annexure-III)
- (e) Commercial Bid (Annexure-IV)
- (f) Compliance Statement (Annexure-V)
- (g) Format of Performance Security (Annexure-VI)
- (h) Declaration (Annexure-VII)
- (i) EMD Details (Annexure-VIII)

3.2. The bidder is expected to examine all instructions, forms, terms and conditions in the Tender Document. In the event of discovery of any missing pages, the bidder shall inform the same to the Section/ Department concerned. Failure to furnish the information required by the Tender Document or submission of a tender not substantially responsive to the Tender Document in every respect will be at the bidder's risk and may result in rejection of the bid.

3.3. The bidder shall not make or cause to be made any alteration, erasure or obliteration to the text of the Tender Document.

## **4. LANGUAGE/FORMAT/SIGNING OF THE BID**

4.1. The bid prepared by the Bidder and all correspondence and documents related to the tender exchanged by the Bidder and IITPKD shall be in English and the Contract shall be construed and interpreted in accordance with that language. If any of the brochures, leaflets or communication is prepared in any language other than English, a translation of such document, correspondence or communication shall also be provided at the cost and risk of the bidder. The translation so provided shall prevail in matters of interpretation. The bidder, with respect to such documents, correspondence and communications, shall bear the costs and risks of such translation.

4.2. The documents comprising the bid shall be typed or written in indelible ink and all the pages shall be signed by the bidder or a person or persons authorized by the bidder. All the pages of the bid shall be numbered and except for unamendable printed, shall be signed by the person or persons authorized.

4.3. The bid shall not contain any internalizations, erasures, overwriting, except to correct errors made by the bidder, in which case the person or persons signing the bid shall initial such corrections with date.

## **5. DOCUMENTS COMPRISING THE BID**

5.1 The Technical and the Commercial Bids shall be sealed by the bidder in separate covers duly superscribed as **Supply, Installation, Testing and Commissioning of Thermo Vacuum Chamber** and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly superscribed as **Supply, Installation, Testing and Commissioning Thermo Vacuum Chamber - Tender No. IITPKD/MECH/SC/036/2019-20**. The technical details of the equipment should be kept inside the Technical Bid envelope and sealed. If the bigger cover is not sealed and marked as required above, IITPKD will assume no responsibility for the bid's misplacement or premature opening.

5.2. Bids must either be spiral bound/stapled together. No loose sheets will be accepted. All pages must be serially numbered.

5.3. The bidder shall furnish, as part of the bid, an EMD (Earnest Money Deposit) @ 2% of the estimated value, as mentioned in the **Annexure-VIII**, in the form of an Account Payee DD/Fixed Deposit Receipt/Bank Guarantee from a nationalised bank in an appropriate format, in favour of **INDIAN INSTITUTE OF TECHNOLOGY PALAKKAD PAYABLE AT PALAKKAD** along with the Tender Document. EMD shall be put in a separate sealed cover and duly superscribed (as was done for the technical/commercial bids). **It should not be enclosed either with Technical Bid or with Financial Bid. It shall be ensured by the supplier that the EMD amount is exactly as mentioned in the Annexure-VIII and NOT on the actual price quoted. In case of the EMD amount being LESS than what is specified in the Annexure-VIII, the bid shall be summarily rejected by the Institute.**

5.3.1. Bids not accompanied by the EMD shall be rejected/disqualified.

5.3.2. Photo/FAX copies of the Demand Draft/Banker's pay orders will not be accepted.

5.3.3. No interest will be paid for the EMD.

5.3.4. EMD of the unsuccessful bidders will be returned to them at the earliest after expiry of the final bid validity and latest by the 30<sup>th</sup> day after the award of the contract.

5.3.5. EMD of the successful bidder shall be returned on receipt of the prescribed Performance Security and after signing of the contract agreement.

5.3.6. EMD shall be forfeited, if the bidder withdraws his bid during the period of validity of the tender.

5.3.7. EMD shall be forfeited, if the successful bidder refuses or neglects to execute the contract or fails to furnish the required Performance Security within the time frame specified by the Institute.

5.4. Documents establishing conformity of the terms and conditions of the Tender Document shall be provided along with the bid. The offer/bids should be sent only for a system or equipment that is available in the market and supplied to a number of customers. A list of customers in India and abroad with details must accompany the quotations. Quotations for a prototype machine will not be accepted.

5.5. Compliance or Confirmation report with reference to the specifications and other terms and conditions should also be obtained from the principal.

5.6. Information related to the agency/bidder such as photocopies of the Registration/PAN/GST/TIN shall be furnished.

5.7. The technical bid should consist of all technical details along with commercial terms and conditions. No prices should be included in technical bid. Financial Bid should indicate item-wise prices for the items mentioned in the technical bid.

5.8. Copy of Tender Document marked "Original" with each page signed and stamped to acknowledge acceptance of the same.

## **6. BID PRICES**

6.1. Prices must be quoted separately for each equipment/item identified.

6.2. Price quoted for equipment/item must include all costs associated with packing, transportation, insurance, delivery of equipment, taxes (separately), loading and unloading on DOOR DELIVERY basis to the institute including its installation, commissioning, integration and validation.

6.3. In case of equipment originating in other countries, prices shall be quoted both on FOB (Port of Shipment) and CIF (Port of Destination) and CIP (Carriage and Insurance Paid). The

comparable prices will be arrived at based on CIP Cochin basis. In the case equipment originating in other countries, the bidder shall provide the following at the time of supply, within 24 hours of despatch:

- a) Supplier's Invoice giving full details of the goods including quantity, value, etc.;
- b) Packing list;
- c) Certificate of country of origin;
- d) Manufacturer's guarantee and Inspection certificate;
- e) Inspection certificate issued by the Purchaser's Inspector;
- f) Insurance Certificate;
- g) Name of the Vessel/Carrier;
- h) Bill of Lading/Airway Bill;
- i) Port of Loading;
- j) Date of Shipment;
- k) Port of Discharge & expected date of arrival of goods and
- l) Any other document(s) as and if required in terms of the contract.

6.4. Prices quoted by the bidder shall be fixed during the validity of the bid.

**6.5 Bids quoted with the terms such as Exworks/FOB/FCA/C&F/CPT etc., will not be accepted and summarily rejected**

**6.6. Insurance Coverage for the equipment for one year against fire / theft / burglary shall be provided by the supplier against the purchase order after the supply, installation, commissioning of the equipment (from the Date of Installation).**

## **7. BID CURRENCY**

7.1. Prices of indigenous equipment/items shall be quoted in Indian Rupees.

7.2. Prices of equipment/items originating in other countries shall be quoted in the currency of country of origin and the portion of allied work and services, which are to be undertaken in India, are to be quoted in the Indian Currency. The comparison of financial bids would be done after converting the currency value in INR based on RBI rates applicable on the date of opening of the tender.

## **8. CONFORMITY OF THE TENDER DOCUMENT**

8.1. The Bidder shall furnish, as part of its bid, documents establishing the conformity of the Equipment that the Bidder proposes to supply under the Contract to the requirements of the Purchaser, as given in the Tender Document.

8.2. The documentary evidence of conformity of the Equipment to the Tender Document may be in the form of written descriptions supported by literature/diagrams/certifications, including:

- (a) A detailed description of the essential technical, functional and performance characteristics of the equipment that the Bidder is proposing to supply;
- (b) Technical details of the major subsystems/components of the equipment;

## **9. PERIOD OF VALIDITY OF BIDS**

Bids shall remain valid for a period of 180 days after the date of deadline for submission of bids prescribed by the Purchaser.

## **10. MODIFICATION AND WITHDRAWAL OF BIDS**

10.1. The Bidder may modify or withdraw the bid after submission, provided a written notice of the modification or withdrawal is received by the Purchaser prior to the deadline prescribed for bid submission.

10.2. The Bidder's modifications shall be prepared, sealed, marked and dispatched in an envelope duly marked BID MODIFICATION.

10.3. A Bidder wishing to withdraw the bid shall notify the Purchaser in writing prior to the deadline prescribed for bid submission. The withdrawal notice shall:

- (a) be addressed to the Purchaser at the specified address and
- (b) bear the reference number and the title of the project, and the words BID WITHDRAWAL NOTICE. Bid withdrawal notices received after the bid submission deadline will be ignored and the bid submitted prior to that will be deemed to be a valid bid.

10.4. No Bid may be modified subsequent to the deadline for submission of Bids.

10.5. No bid may be withdrawn in the interval between the bid submission deadline and the expiration of the bid validity period. Withdrawal of a bid during this interval may result in the forfeiture of the Bidder's EMD.

## **11. OPENING AND EXAMINATION OF BIDS**

11.1. The Technical bids will be opened on the prescribed date and time as mentioned in the Bid document. Bidders or their representative may be present during the opening of technical bid, if they wish to be present.

11.2. The purchaser will evaluate the technical bids. Those bids, whose technical bids fulfill the technical requirements and responsive to the tender requirements will be considered. Those bids which found to be either non-responsive, not satisfying the technical requirements or both will not be considered and will be rejected.

11.3. The Price bids of the successful bidders on the basis of evaluation as mentioned in will be considered for the next stage for opening.

11.4. The Purchaser will examine the bids to determine whether they are complete, whether any computational errors have been made, whether required security has been furnished, whether the documents have been properly signed and whether the bids are generally in order.

11.5. Arithmetical errors will be rectified on the following basis. If there is a discrepancy between the unit price and the total price, which is obtained by multiplying the unit price and quantity, or between subtotals and the total price, the unit or subtotal price shall prevail and the total price shall be corrected. If there is a discrepancy between words and figures, the amount in words shall prevail. If a Bidder does not accept the correction of errors, the bid will be rejected and its EMD may be forfeited.

11.6. The Purchaser may waive any minor non-conformity or irregularity in a bid that does not constitute a material deviation, provided such waiver does not prejudice or affect the relative ranking of any Bidder.

11.7. Prior to the detailed evaluation, the Purchaser will determine whether each bid is complete and is substantially responsive to the Tender Document. For purposes of this determination, a substantially responsive bid is one that conforms to all the terms, conditions and specifications of the Tender Document without material deviations, exceptions, objections, conditionality or reservations. A material deviation, exception, objection, conditionality, or reservation is:

(a) One that limits in any substantial way the scope, quality, or performance of the Equipment; **OR**

(b) One that limits, in any substantial way that is inconsistent with the Tender Document, the Purchaser's rights or the successful Bidder's obligations under the Contract: and

(c) One that the acceptance of which would unfairly affect the competitive position of other Bidders who have submitted substantially responsive bids.

11.8. If a bid is not substantially responsive, it shall be rejected by the Purchaser and may not subsequently be made responsive by the Bidder by correction of the non-conformity.

The Purchaser's determination of bid responsiveness will be based on the contents of the bid itself and any written clarifications submitted by the Bidder.

## **12. CLARIFICATION OF BIDS**

During the bid evaluation, the Purchaser may, at its discretion, ask the Bidder for a clarification of its bid. The request for clarification and the response shall be in writing, and no change in the price or substance of the bid shall be sought, offered or permitted.

## **13. EVALUATION OF RESPONSIVE BIDS**

The Purchaser will evaluate the bids that have been determined to be substantially responsive.

## **14. CONTACTING THE PURCHASER**

14.1. From the time of bid opening to the time of Contract award, if any Bidder wishes to contact the Purchaser on any matter related to the bid, it shall do so in writing.

14.2. If a Bidder tries to directly influence the Purchaser or otherwise interfere in the bid evaluation process and the Contract award decision, his bid shall be rejected.

#### **15. AWARD CRITERIA**

15.1 Purchaser will award the Contract to the Bidder, whose bid has been determined to be substantially responsive and evaluated as the lowest quote.

15.2. The Institute reserves the right to buy different items/quantity from different bidders considering price of individual/group of equipment or any other factors as decided by the Committee.

#### **16. PURCHASER'S RIGHT TO ACCEPT/REJECT BIDS**

16.1. The Purchaser reserves the right to accept or reject any bid or to annul the bidding process and reject all bids at any time prior to Contract award, without thereby incurring any liability to the Bidders.

16.2. The Purchaser reserves the right to negotiate with the Bidder, whose bid has been evaluated as the lowest quote.

#### **17. AWARD OF PURCHASE ORDER**

17.1. Prior to the expiration of the period of bid validity, the Purchaser will issue the Letter of Intent / Purchase Order to the successful Bidder in writing.

17.2 The Purchase Order will constitute the foundation of the Contract.

17.3. Upon the successful Bidder's furnishing of the copy of the Purchase Order duly signed on each page and the Performance Security, for the equipment ordered in foreign currency, the Purchaser will open a letter of credit (LC) in a convenient Nationalized Bank in India. For opening of LC necessary arrangements shall be provided by the supplier or its authorized agents.

#### **18. CONTRACT AGREEMENT**

18.1. Within fifteen (15) days of receipt of the Purchase Order, the successful Bidder shall sign and date its copy on each page and return it to the Purchaser, along with the Performance Security.

18.2. Copy of Purchase Order duly signed and dated by the successful Bidder on each page shall constitute the Contract Agreement.

#### **19. PERFORMANCE SECURITY**

**19.1. The performance security shall be submitted within FIFTEEN (15) days of receipt of the material by the purchaser. The successful bidder shall furnish the Performance Security equal to 5% of the order / contract value (excluding the value of annual maintenance charges). The Performance Security shall be valid all along the warranty period and shall extend upto sixty (60) days after the date of completion of warranty period. It shall be ensured by the successful bidder that the validity of the Performance Security submitted is extended depending on the date of commencement of the Warranty.**

19.2. The performance security shall be a bank guarantee (in the format as provided in **Annexure-VI** of the bidding documents) issued by the Indian Scheduled bank acceptable to the Purchaser or a Demand Draft favouring, **INDIAN INSTITUTE OF TECHNOLOGY PALAKKAD** payable at PALAKKAD.

19.3. In case the successful bidder is a foreign company and wishes to submit Performance Security in the form of Bank Guarantee, the Bank Guarantee should be routed through the Beneficiary Bank to the end user bank. Otherwise, the Indian Agent of the foreign vendor shall submit a Bank Guarantee from a Nationalized Bank of India. The following documents shall be submitted in case of an Indian agent submitting the Performance Security on behalf of his principal:

- Foreign principal's proforma invoice indicating the commission payable to the Indian agent and nature of after-sales service to be rendered by the Indian agent.
- Copy of the agency agreement with the foreign principal and the precise relationship between them and their mutual interest in the business.

19.4. The performance security shall automatically become null and void once all the obligations of the Supplier under the Contract have been fulfilled, including, but not limited

to, any obligations during the Warranty Period and any extensions to the period. The performance security shall be returned to the Supplier not later than fifteen (15) days after its expiration.

19.5. Failure of the successful Bidder to comply with the requirements shall constitute enough grounds for the annulment of the award and forfeiture of the EMD, in which event the Purchaser may make the award to the next lowest evaluated bid submitted by a qualified Bidder or call for new bids.

## **20. CONTRACT DOCUMENTS**

20.1. All documents forming part of the Contract (and all parts of these documents) are intended to be correlative, complementary and mutually explanatory. The Contract shall be read as a whole.

20.2. The order of precedence of the Contract documents shall be as follows:

- (i) Contract Agreement
- (ii) All other Forms
- (iii) Equipment and their requirements
- (iv) Supplier's Bid
- (v) Tender Document

## **21. AMENDMENT TO CONTRACT**

No amendment or other variation of the Contract shall be effective unless it is in writing, is dated, expressly refers to the Contract and is signed by a duly authorized representative of each party to the Contract.

## **22. SUPPLIER'S RESPONSIBILITIES**

22.1. The Supplier's obligations involve:

- (a) Supply of equipment/items given in Tender Document.
- (b) Making operational, the equipment (installation, commissioning, testing and validation of the equipment).
- (c) Development of test methods and applications.
- (d) Training, at the cost of Supplier, of personnel in operation, day-to-day maintenance and troubleshooting of the Equipment.
- (e) Supply of Material (instruction/operation/service/maintenance manuals including drawings & circuit diagrams and application notes), Calibration Certificates (where applicable, traceable to national/international standards) and any other documents specified in the Contract.

22.2. The Supplier shall, unless specifically excluded in the Contract, perform all such work and/or supply all such items, services and materials not specifically mentioned in the Contract but that can be reasonably inferred from the Contract as being required for installation, commissioning, integration and validation of equipment as if such work and/or items and materials were expressly mentioned in the Contract.

22.3. The Supplier shall comply with all laws in force in India. The laws will include all national, provincial, municipal or other laws that affect the performance of the Contract and are binding upon the Supplier. The Supplier shall indemnify and hold harmless, the Purchaser from and against any and all liabilities, damages, claims, fines, penalties and expenses of whatever nature, arising or resulting from the violation of such laws by the Supplier.

## **23. TIME FOR SUPPLY, INSTALLATION, COMMISSIONING AND VALIDATION OF THE EQUIPMENT**

23.1. The Supplier shall supply the equipment within the period specified in the tender document i.e. within **SIX (6) Months** of signing the purchase order or within the period mutually agreed between purchaser and supplier.

23.2. The Supplier shall thereafter proceed with the installation, commissioning, integration and validation and demonstrate operational acceptance of the equipment within the period specified. The equipment shall be installed and commissioned by the successful bidder within 20 to 25 days from the date of its receipt.

23.3. The tenderer should indicate clearly the time required for delivery of the item. In case there is any deviation in the delivery schedule, liquidated damages clause will be enforced or penalty for the delayed supply period will be levied.



23.4. In the event of failure of supply of the item/equipment within the stipulated delivery schedule, the Purchaser has all the right to purchase the item/equipment from other sources on the total risk of the Supplier under the risk purchase clause.

## **24. TERMS OF PAYMENT**

24.1. No Advance payment will be made for indigenous purchase. For indigenous equipment, 100% payment will be released after satisfactory delivery, acceptance, installation, commissioning, integration and validation of the equipment and against the installation report/certification provided jointly by the supplier (technical representative involved in the process of installation) and the faculty in-charge of the user department/facility.

24.2. For imported equipment, a Letter of Credit (LC) shall be opened for 100% CIP price on receipt of the acknowledgement of the purchase order. However, 90% of the LC amount only shall be released on proof of the shipment of the consignment with necessary documents to be provided in detail at the time of placing of the purchase order. Balance 10% of the LC amount shall be released upon the receipt of a performance security of 5% of the total value of the purchase order, as per the clause 19 above and installation, commissioning, integration, validation and installation report/certification jointly given by the end user and the supplier.

**24.3. For equipment ordered in foreign currency, opening of LC, payment of Customs Duty and clearance of goods shall be done by the supplier or its authorized Indian agent. The custom duty as applicable after considering eligible concessions based on DSIR exemption etc. will only be paid by the purchaser. A copy of the DSIR Certificate shall be provided by the purchaser upon request, for customs duty exemption.**

24.4. Payment for annual maintenance contract (AMC) shall be released annually at the end of each year of the AMC Period.

24.5. If any time before the delivery of the equipment, it is found that the same equipment had been offered to another party in India at a lower rate, payment shall be restricted to the extent of such lower rate and the Supplier shall be liable to pay the Purchaser the difference in two rates i.e. excess charged over such lower rate, if payment had been made by the purchaser. The purchaser will look into a reasonable past period to ensure this.

## **25. TAXES AND DUTIES**

The Supplier should ensure payment of all taxes, duties, levies and charges assessed by all municipal, state or national government authorities, in connection with the Goods and Services supplied under the Contract.

## **26. PRODUCT UPGRADES**

The Supplier shall continue to support and maintain the version/model of the Equipment supplied by upgrading the software and the hardware as and when amendments are carried out in the existing version or the product is upgraded. Whereas upgrades to the software shall be supplied free of cost, the Supplier may charge for upgrade in hardware provided it is of major nature. An upgraded higher version of the instrument and software related with the instrument shall be supplied.

## **27. PENALTIES**

27.1. If the Supplier fails to complete any of the activities in accordance with the time specified for it, or any extension of the time granted by the Purchaser, the Supplier shall pay to the Purchaser, penalties at the rate specified in the Tender document.

27.2. The Purchaser reserves the right to terminate the contract if the Supplier defaults on any of the time limits by more than FOUR weeks.

## **28. DEFECT LIABILITY**

28.1. The Supplier warrants that the equipment, including all subassemblies and components provided, shall be free from defects in the design, engineering/manufacturing, workmanship and performance that prevent the Equipment and/or any of its subassemblies and components from fulfilling the equipment requirements or that limit in a material fashion the operation, reliability, accuracy, sensitivity and precision of the Equipment, its subassemblies and

components. Commercial warranty provisions of products supplied under the Contract shall apply to the extent that they do not conflict with the provisions of this Contract.

**28.2. The warranty period shall commence from the date of validation/installation of the equipment and shall extend for the length of time specified in the tender document supra.**

28.3. If during the warranty period any defect is found in the equipment, the Supplier shall promptly, at its sole cost, repair or otherwise make good such defect as well as any damage to the equipment caused by such defect. Any defective equipment, subassembly or component that has been replaced by the Supplier shall become the property of the Supplier and the new substituted/replaced equipment in good condition shall become the property of the purchaser.

28.4. Validation of the equipment shall be carried out by the Supplier each time a major repair is carried out in the equipment during the warranty period.

28.5. Response time for attending to defects shall be 24 to 48 hours, after they are reported to the Supplier or its designated service agent. If the equipment cannot be used for more than TWO working days by reason of such defect and/or making good of such defect, the warranty period for the equipment shall be extended by a period equal to the period during which the equipment could not be used by the Purchaser because of such defect and/or making good of such defect.

## **29. INTELLECTUAL PROPERTY RIGHTS, WARRANTY AND INDEMNITY**

29.1. The Supplier hereby represents and warrants that the equipment as supplied, installed and commissioned along with its application software and copying of manuals and other documents provided to the Purchaser in accordance with the Contract does not and shall not infringe any Intellectual Property Rights held by any third party.

29.2. The Supplier shall indemnify and hold harmless the Purchaser from and against any and/or all losses, liabilities and costs (including losses, liabilities and costs incurred in defending a claim alleging such a liability), that the Purchaser may suffer because of any infringement or alleged infringement of any Intellectual Property Rights.

29.3. The offer should clearly specify the warranty or guarantee period for the machinery/equipment. Any extended warranty offered for the same shall be mentioned separately.

## **30. EFFECT OF FORCE MAJEURE**

30.1. If the Supplier is prevented, hindered, or delayed from or in performing any of its obligations under the Contract by an event of Force Majeure, then it shall notify the Purchaser in writing of the occurrence of such event and the circumstances of the event of Force Majeure within fifteen (15) days after the occurrence of such event.

30.2. The Supplier, when affected by the event of Force Majeure shall use reasonable efforts to mitigate the effect of the event of Force Majeure upon its performance of the Contract and to fulfill its obligations under the Contract, but without prejudice to Purchaser's right to terminate the Contract.

30.3. No delay or non-performance by the Supplier caused by the occurrence of any event of Force Majeure shall:

- (a) Constitute a default or breach of the Contract;
- (b) Give rise to any claim for damages or additional cost or expense occasioned by the delay or non-performance.

30.4. If the performance of the Contract is substantially prevented, hindered, or delayed for a single period of more than THIRTY days or an aggregate period of more than SIXTY days on account of one or more events of Force Majeure, the Purchaser shall have the right to terminate the Contract by giving a notice to the Supplier.

## **31. EXTENSION OF TIME LIMITS FOR SUPPLY AND MAKING OPERATIONAL, THE EQUIPMENT**

31.1. The time limit for supply, installation & commissioning, integration & validation shall be extended if the supply is delayed or impeded in the performance of any of its obligations under the Contract by reason of any of the following:

- (a) Any occurrence of Force Majeure;
- (b) Any other matter specifically mentioned in the Contract;

By such period as shall be fair and reasonable in all the circumstances and as shall fairly reflect the delay or impediment sustained by the Supplier.

**31.2. The Supplier shall assist the institute in assembling/disassembling the Equipment at free of cost, while it is relocated between the campuses, within the warranty period.**

**32. ASSIGNMENT**

The Supplier shall not, without the prior written consent of the Purchaser, assign to any third party, the Contract or any part thereof.

**33. GOVERNING LAW**

The Contract shall be governed by and interpreted in accordance with the laws of India.

**34. SETTLEMENT OF DISPUTES**

Any dispute or claim arising out of/relating to this Contract or the breach, termination or the invalidity thereof, shall be settled by the Hon'ble Courts of Justice at Palakkad.

**35.** The page number should be marked in all pages serially (including all supporting documents enclosed with the tender document) and the declaration for the same shall be submitted by the bidder as in **Annexure-VII**.

**36.** IITPKD reserves the right to accept or reject any or all the tenders in part or whole or may cancel the tender at its sole discretion without assigning any reason whatsoever. No further correspondence in this regard will be entertained.

**REGISTRAR**

**TECHNICAL SPECIFICATIONS****Equipment Name: Thermo Vacuum Chamber**

Sl. No.	Parameter	Specification
<b>1. Thermo-vacuum chamber</b>		
1.1	Shape of construction	Cylindrical
1.2	Material of construction	SS 304 (Stainless steel 304)
1.3	Inner Diameter	700 mm
1.4	Cylindrical length	600 mm
1.5	Front door	Strong hinge supported dished door
1.6	Vacuum sealing	All vacuum sealing between door, ports, feedthroughs or any other openings and chamber flanges should be through <b>jointless</b> Viton 'O' ring
1.7	Chamber rear end	Welded dished end
<b>2. Platform for supporting job</b>		
2.1	Location	Inside the vacuum chamber located at 100 mm from the bottom of the chamber
2.2	Dimensions	Width = 490 mm depth(length) = 490 mm
2.3	Material of construction	SS 304 (Stainless steel 304)
2.4	Type of mounting	Sliding platform mounted on wheels supported on welded guide rails with stopper arrangement. It should be possible to slide the platform on the rails and remove it from the chamber, if required by the operator.
2.5	Weight carrying capacity	The platform should be capable of carrying a job gross weight of 300 kilograms.
2.6	Cooling/heating/temperature control requirement for the platform	Not required
<b>3. Chamber Ports</b>		
3.1	Dimension	Inner diameter (ID) 150 mm
3.2	Number	Five (5)
3.3	Location	Center of top, bottom, left, right chamber walls and rear dished end

3.4	Port dummy flanges	All ports should be provided with dummy flanges.
3.5	Additional small ports on dummy flanges	The dummy flanges fixed on the left and right walls and rear dished end wall of the chamber should have <b>one (1) number each of KF 25 and KF 50 ports with respective dummy flanges</b>
<b>Note : The port fixed at the bottom wall of the chamber should be used for evacuation (fixing of the vacuum pumps).</b>		
3.6	Additional port for liquid feed through  (This feedthrough is required to supply cooling fluid to a heat exchanger plate kept on the job platform inside the chamber.If required, feedthrough drawings shall be provided by customer during design stage, else the manufacturer can choose to provide their own drawing to be approved by the customer).  <b>Note: Scope is supply is limited to the feedthroughs only.</b>	One (1) number KF 50 port with liquid feed through and dummy should be provided on the right wall of the chamber next to the ID 150 mm port as mentioned earlier.  The feedthrough should consist of half inch (1/2”) SS 304/316 or Copper inlet-outlet tubing with suitable ferrule based terminations on the vacuum as well as atmosphere sides.
3.7	Viewport	One (1) number KF 100 viewport with glass window should be provided at the center of the front hinged dished door.
<b>4.0 Chamber support structure and frame</b>		
4.1	Support structure	MS (Mild Steel) frame with castor wheels and locking arrangement.
<b>Note: The support structure should house the vacuum pumping system, related equipment and control console as well.</b>		
<b>5. Vacuum Pumping System</b>		
5.1	Diffusion Pump with cryo cooled trap to prevent oil back streaming and facilitate clean vacuum	700 litres/min
5.2	Rotary vacuum pump	500 litres/min
5.3	Roots vacuum pump	250 m <sup>3</sup> /hr
<b>Note: The pump capacities are indicative only. Pumps may be chosen suitably by the manufacturer to meet the ultimate vacuum level mentioned in Sl. No. 5.5.</b>		
5.4	Valves for high vacuum, roughing, backing and venting and operations	Manual right angle valves
5.5	Ultimate vacuum level to be achieved	1 × 10 <sup>-6</sup> millibar range
5.6	Air admittance valve	Manual
5.7	Gas line and valve	A gas inlet/outlet line with needle valve should be provided to fill the chamber with an inert gas for experiments and

		leakage checking, if required. Manually operated valves should be used.
<b>6. Vacuum measurement system</b>		
6.1	Vacuum measurement range	One atmosphere to $10^{-6}$ millibar range
6.2	Pirani gauges with control unit	2 numbers
6.3	Penning gauge with control unit	1 number
<b>7. Leak rate</b>		
7.1	Clean, dry empty chamber after thorough degassing	Not to exceed $1 \times 10^{-3}$ m. litres/sec
7.2	Leak checking	To be done using a Mass Spectrometer Leak Detector (MSLD) having a sensitivity of $1 \times 10^{-8}$ milli bar litres/sec
<b>8. Thermal Shroud</b>		
8.1	Material of construction	Electrolytic copper or stainless steel
8.2	Location of shrouds	Shrouds are housed inside chamber
8.3	Number of shrouds	Three (3)  i) One shroud for the cylindrical portion (inner diameter 650 mm and length 600 mm)  ii) Two shrouds to be provided on the front and rear dished ends
8.4	Curvature of shrouds	Should closely match the inner curvature of the chamber
8.5	Temperature uniformity and control	All shrouds should be controlled using a single thermal circulator to maintain uniform temperature
8.6	Shroud movement	Shrouds should be detachable in nature and fixed on rails to permit axial movement
8.7	Painting of shroud	Inner surface of the shrouds should be painted using vacuum compatible epoxy black (high emissivity) paint.
8.8	Normal temperature range of operation of the shroud (thermal circulator should be provided for this temperature range)	-40 to 100 °C
8.9	Design temperature range of operation of the shroud (shroud tubing and brazing should be designed to withstand thermal cycling in the mentioned temperature range)	-196 to 200 °C
8.10	Temperature uniformity on the shroud	better than or equal to $\pm 1$ °C at steady state

8.11	Shroud temperature measurement and monitoring	Suitable temperature measurement sensors and control instrumentation should be provided. Temperature readout display should be provided in the control console.
8.12	Cut outs in shroud	Suitable cutouts should be provided in the shroud to facilitate passage of instrumentation cables and fluid lines. These cables and lines shall enter the chamber through the KF ports provided on the 150 mm inner diameter flanges (as mentioned earlier in Sl. No. 3) and terminate at the job bearing platform.
<b>9. Thermal Circulator</b>		
9.1	Temperature range	-40 to 100 °C
9.2	Temperature stability	± 0.05 °C
9.3	Display resolution	± 0.01 °C
9.4	Cooling power (approximately calculated based on heat dissipation from the job and power required to cool the shroud. Exact values may be decided during technical design finalization stage. However, major deviations from the mentioned values are not expected)	2000 watts at 20 °C 1500 watts at 0 °C
<b>10. Control console</b>		
10.1	Control console	Suitable control console to readout relevant parameters and control the operation of the thermo-vacuum chamber should be provided and should be preferably located with in the supporting frame structure
<b>11. Safety Devices</b>		
11.1	Safety and protection devices and interlocks	Suitable safety devices and interlocks should be provided wherever necessary to prevent damage to the system and operator due to machine malfunction or operation error
<b>12. Other requirements</b>		
12.1	Delivery time	With in six months of technically and commercially clear purchase order
12.2	Onsite Warranty	1 year and above

**PRE-QUALIFICATION CRITERIA FOR BIDDERS**

Only those bidders fulfilling the following criteria should respond to the tender.

1. Bidder should be either an Original Equipment Manufacturer (OEM) or designer or authorized distributor of an OEM of **Thermo Vacuum Chamber** allied field.
2. The bidder should be a company registered under the Companies Act, 1956/2013 OR a Limited Liability Partnership / a registered partnership firm OR a sole-proprietorship entity. Appropriate Registration incorporation certificate must be submitted.
3. The bidder should have presence in the business of design, development, supply and installation of standard as well as customized vacuum technology equipment in Indian as well as global market for at least past 20 years.
4. The bidder should have ISO 9001:2015 certification and preferably ISO 14001:2015 and OSHAS 18001:2007 certifications.
5. The bidder should preferably adhere to NADCAP (National Aerospace and Defense Contractors Accreditation Program) standards and should provide necessary evidence of the same .
6. Bidders, who are bidding for this NIT shall,
  - (i). Have documentary evidence (Purchase Order) of having completed **Supply, Installation, Testing and Commissioning of at least 1 unit of Thermo Vacuum Chamber** to a Centrally Funded Technical Institution/Research laboratory (e.g., IIT, NIT, IISc, IISER, ISRO, DRDO etc.). The bidder must provide a certificate of satisfactory performance of the supplied equipment from the institute/organization to which they have recently supplied. Contact details of the faculty In-charge of the installed setup must be provided.
  - (ii). Have an Average Annual Turnover of atleast **Rs. 2.5 Crore** during each of the last three financial years (2015-16, 2016-17, 2017-18). The bidder shall enclose the audited statements of the indicated financial years, which should have been certified by a Chartered Accountant.
7. The bidder must have a registered office and/or service center in Karnataka/Tamil Nadu/Telangana/Andhra Pradesh/Maharashtra or Kerala. Certificate of registration for the offices to be provided. Details about scope of service activities provided by the service centres must be provided. The contact details of the service engineers must be provided.
8. Preference will be given to bidders who have authorized service center and personnel located in the State of Kerala, India. Details of such service centers such as address and name and contact details of service personnel should be provided.
9. The bidder must provide detailed specification of each equipment/item. Model numbers, data sheets and brochures must be included for each quoted equipment/accessories/item. If not, bid can be disqualified on technical grounds.
10. The Institute reserves the right ask for photographs/CAD drawings/ design proofs to satisfy themselves of the proven capabilities of the system being offered. The bidder must provide these details with in two working days of receiving such a request via email. Decision regarding technical compliance of the bidder can be taken on the basis of this information.
11. Compliance sheet for the technical specification and OEM Brochure have to be attached along with the Technical bid. Vendor has to fill the compliance sheet and mention page number or reference number in OEM brochure. Unfilled / partially filled sheets lead to disqualification.



**TECHNO-COMMERCIAL PARAMETERS****(To be enclosed in a separate sealed cover with the Technical Bid)****PART-I****1. Company Profile**

- a) Name
- b) Postal address of the registered office
- c) Name & Designation of CEO / Director
- d) Nature of Business (Proprietary / Partnership/ any other mode)
- e) Email ID & Contact number(s) of CEO / Director
- f) No. of years of operations in India
- g) Year of Establishment
- h) Location of offices in India / abroad

**2. Alliances for the purpose of this Bid**

- a) Details of alliance(s)

- b) Type of alliance(s)

**3. Experience/Credentials**

- a) No. of similar units installed in India

- b) No. of similar units installed in Karnataka/Tamil Nadu/ Telangana/Andhra Pradesh or Kerala or any other nearby city/town

- c) List of satisfied clients in India

- d) Testimonials from atleast three clients shall be attached

**4. Service Support in India**

- a) Track record of service provided with supporting documents during last 3 years
- b) Location of service centers

- c) Number of trained service engineers

- d) Number of trained service engineers exclusively dedicated to each equipment offered

- e) Number of trained service engineers for the equipments offered stationed in Karnataka/Tamil Nadu/ Telangana/Andhra Pradesh or Kerala

- f) Number of application specialists

- g) Whether the OEM makes available any service

- h) Support in India

**5. Availability of spares in India**

- a) Whether the service set up maintains stock of essential spares in India

- b) Lead time for supply of essential spares

## **PART-II**

### **1. Name of the Bidder**

- a) Postal address
- b) Telex / Fax number
- c) Telephone (Landline)
- d) Mobile No.:
- e) Email address:
- f) Type of firm: Propriety/ Private/ Private Ltd/ MNC/ Cooperative /Govt. undertaking
- g) Name of the Proprietor /Partners
- h) Registration No.
- i) Year of commencement of Manufacturing the equipment
- j) PAN
- k) TIN
- l) GST Number

### **2. Details of EMD of 2% of estimated value as mentioned in Annexure-VII**

DD No.: \_\_\_\_\_ Date: \_\_\_\_\_ Amount: \_\_\_\_\_ Bank/Branch: \_\_\_\_\_

### **3. Total Annual Turnover (2015-16, 2016-17 and 2017-18)**

(Documents shall be certified by Competent Authority)

### **4. Has the firm ever been debarred/blacklisted by any Govt. Organization/Dept.? If 'yes' the details thereof.**

**Signature** \_\_\_\_\_

**Name** \_\_\_\_\_

**Designation** \_\_\_\_\_

**Date:** \_\_\_\_\_ **Place:** \_\_\_\_\_

**Seal of Company** \_\_\_\_\_

### **Note:**

1. Compliance Statement to specifications of the equipment to be provided by the tenderer as in **Annexure-IV**.
2. All equipment must operate at 230V/50 Hz single phase and/or equivalent three phase electrical power.
3. Quoted model shall be in accordance to the geographical location.
4. All the above details shall be related to the vendor for the items quoted.

**Signature and Seal of the Bidder**

**COMMERCIAL BID**  
(To be enclosed in separate sealed cover)

**THE PRICES SHALL NOT BE MENTIONED HERE WHEN SUBMITTED AS A PART OF THE TECHNICAL BID, AS PER THE CLAUSE NO.5.8**

1. The price of the equipment, whose details are available in the **Annexure-I** is to be given in the format mentioned below:

Sl. No.	Name of the Equipment	No. of Units/ Sets	Cost per Unit/Set	Total Cost
1	<b>Supply, Installation, Testing and Commissioning of Thermo Vacuum Chamber</b>	<b>1 No</b>		
<b>Total Amount</b>				
<b>Taxes : GST @ 5%</b>				
<b>Grand Total</b>				
<b>Amount in Words:</b>				

2. The quote should include a warranty of **ONE YEAR** from the date of commissioning/installation of the equipment and AMC for **FIVE YEARS**. The format for AMC is provided below:

Sl. No.	Annual Maintenance Contract (AMC) after the Warranty Period				
Name of the Equipment	2 <sup>ND</sup> YEAR	3 <sup>RD</sup> YEAR	4 <sup>TH</sup> YEAR	5 <sup>TH</sup> YEAR	6 <sup>TH</sup> YEAR

**Note:**

- a) Rates quoted shall be exclusive of taxes and duties.
- b) Taxes and duty component should be mentioned separately.
- c) Bids quoted with the terms such as Exworks/FOB/FCA/C&F/CPT etc., will not be accepted and summarily rejected**
- d) Maximum educational discount as could be offered should be mentioned.
- e) Price quoted for equipment must include all costs associated with packing, transportation, Insurance, delivery of equipment, loading and unloading on **DOOR DELIVERY** basis to the Transit Campus of **IIT Palakkad** including its installation, commissioning, integration, testing and validation. Duties and Taxes must be mentioned separately.
- f) Concessional GST@ 5% will be paid extra against GOI Notification No. 45 & 47/2017, dated 14.11.2017 against Proforma Invoice.
- g) TDS @ 2% will be deducted as per CBEC Circular No.65/39/2018-DOR dated 14.09.2018.
- h) Do not quote the optional items or additional items unless otherwise mentioned in the Tender document / specifications.
- i) The institute reserves the right to procure the equipment with or without the AMC quoted. However the rates will be fixed as per the quotation against this tender. **While quoting for the AMC, the supplier should mention its detailed scope including scope of services covered under the AMC.**

**COMPLIANCE STATEMENT**  
**(Part of Technical Bid)**

The vendor shall,

1. Prepare, sign and submit the Compliance Statement of the specification of the equipment in the format given below along with the technical bid in the company letter head.
2. Submit separate Compliance Statement of specification sheets for each item/equipment.
3. Ensure that the component number and heading in the Technical Specifications is clearly mentioned in the document. If there are any deviations from the specifications mentioned by IIT Palakkad, the vendor should clearly indicate the deviations and give reasons for the deviation with proper justification.
4. Provide the technical leaflet/literature/catalogue or any relevant document for all the quoted equipment to IIT Palakkad. The information provided in the compliance statement without supporting documents will not be considered for the evaluation of the technical bid and will be treated as non-compliance and may lead to the disqualification of the technical bid.
5. Clearly respond to every requirement given in the technical specifications. Lack of clarity may be considered as lack of information and may subsequently lead to disqualification of the technical bid.

**Format of Compliance Statement:**

Item No.	IIT Palakkad's technical specification of components as given in Annexure-I	Specifications of model quoted by the vendor	Vendor's specification complies with IIT Palakkad's technical specification? (YES/ NO)	Deviation, if any, to be indicated in unambiguous terms

**FORMAT OF PERFORMANCE SECURITY**

1. This deed of Guarantee made this day of \_\_\_\_\_ between Bank of \_\_\_\_\_ (hereinafter called the "Bank") of the one part, and Indian Institute of Technology Palakkad (hereinafter called "the Purchaser") of the other part.
  2. Whereas the Purchaser has awarded the contract for Supply, Installation, Commissioning, Integration and Validation of \_\_\_\_\_ (name of the equipment) (hereinafter called the contract) to \_\_\_\_\_ (hereinafter called the Supplier); (Name of the Supplier)
  3. AND WHEREAS the Supplier is bound by the said Contract to submit to the Purchaser a Performance Security for a total amount of Rs. \_\_\_\_\_ (Amount in figures and words).
  4. Now, I/we the undersigned, being fully authorized to sign and to incur obligations for and on behalf of and in the name of \_\_\_\_\_ (Full name of Bank), hereby declare that the said Bank will guarantee the Purchaser the full amount of Rs. \_\_\_\_\_ (Amount in figures and words) as stated above.
  5. After the Supplier has signed the aforementioned Contract with the Purchaser, the Bank is engaged to pay the Purchaser, any amount up to and inclusive of the aforementioned full amount upon written order from the Purchaser to indemnify the Purchaser for any liability of damage resulting from any defects or shortcomings of the Supplier under the Contract mentioned above, whether these defects or shortcomings are actual or estimated. The Bank will deliver the money required by the Purchaser immediately on demand without delay without reference to the Supplier and without the necessity of a previous notice or of judicial or administrative procedures and without it being necessary to prove to the Bank the liability or damages resulting from any defects or shortcomings of the Supplier. The Bank shall pay to the Purchaser any money so demanded notwithstanding any dispute/disputes raised by the Supplier in any suit or proceedings pending before any Court relating thereto and the liability under this guarantee shall be absolute and unequivocal.
  6. This Guarantee is valid for a period of thirty six months from the date of signing. (Initial period for which this Guarantee will be valid must be for at least thirty (30) days longer than the anticipated expiry date of warranty period).
  7. At any time during the period in which this Guarantee is still valid, if the Purchaser agrees to grant a time extension to the Supplier or if the Supplier fails to complete the work within the time of completion as stated in the Contract, or fails to discharge himself of the liability or damages as stated under Para 5 above, the Bank shall extend this Guarantee under the same conditions for the required time on demand by the Purchaser and at the cost of the Supplier.
  8. The Guarantee herein before contained shall not be affected by any change in the Constitution of the Bank or of the Supplier.
  9. The neglect or forbearance of the Purchaser in enforcement of payment of any moneys, the payment whereof is intended to be hereby secured or the giving of time by the Purchaser for the payment hereof shall in no way relieve the bank of its liability under this deed.
  10. The expressions "the Purchaser", "the Bank" and "the Supplier" herein before used shall include their respective successors and assigns.
- In witness whereof I/We of the bank have signed and sealed this guarantee on the \_\_\_\_\_ day of \_\_\_\_\_ (Month & Year) being herewith duly authorized.  
 For and on behalf of the \_\_\_\_\_ Bank.

**Signature of Authority**

Name of the Official Name: ..... Designation: .....  
Stamp/Seal of the Bank: .....

Signed, sealed and delivered for and on behalf of the Bank by the above named \_\_\_\_\_ in the presence of:

**Witness 1**

**Witness 2**

Signature .....  
Name .....  
Address .....

Signature .....  
Name .....  
Address .....

**DECLARATION**

We hereby undertake that there are \_\_\_\_\_ pages, serially numbered, in the submitted tender including the supporting documents. (Please serially number all the pages including blank page, if any).

We have submitted our principal's exclusive authorization letter which is specific for this tender No. \_\_\_\_\_ dated \_\_\_\_\_.

**Signature and Seal of the Bidder**

**ANNEXURE-VIII**

**EMD DETAILS**

<b>Sl. No.</b>	<b>Description of Item</b>	<b>Quantity</b>	<b>Estimated Value (INR)</b>	<b>EMD (INR)</b>
<b>1</b>	<b>Supply, Installation, Testing and Commissioning of Thermo Vacuum Chamber</b>	<b>1 No</b>	<b>24 Lakh</b>	<b>48,000/-</b>

**Signature and Seal of the Bidder**