



**CORRIGENDUM - I, Dated 20-09-2024**

**Tender No. TENDER/2024-25/136**  
**Date of Publication: 06-09-2024**  
**Date/Time of Closing: 27-09-2024, 15:00 hours**

Indian Institute of Technology Palakkad Invites Tender under Two-bid system for the:

**SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF  
NANO PARTICLE AND ZETA POTENTIAL ANALYSER**

<b>Sl. No.</b>	<b>Reference</b>	<b>Tender Specification</b>	<b>Corrigendum</b>
1.	Page No. 8 Annexure - I: Sl.No. 3: Measurement Angles	Minimum 2 Measurement Angles between 10 to 90 Degrees.	Minimum 2 Measurement Angles between 10 to 175 Degrees.

**All other Terms and Conditions of the Tender remain unchanged.**

**REGISTRAR**  
**IIT PALAKKAD**